

G35 Datasheet



AQL 1.5
 EN455 Parts 1, 2, 3 & 4
 MDD 93/42/EEC
 EN374
 Class 1 Medical Device



Blue Nitrile Powder Free Gloves

General Details

Material	Acrylonitrile Butadiene (Nitrile), non sterile
Surface Finish	Powder Free
Colour	Blue
Geometry	Ambidextrous
Texture	Finger Textured
Cuff and Finishing	Beaded

Glove Specifications

Size	Length mm	Palm Width mm	Thickness		
			Cuff	Palm	Finger
XS	240	75 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
S	240	80 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
M	240	95 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
L	240	110 ± 10	Min. 0.05	Min. 0.05	Min. 0.05
XL	240	≥ 110	Min. 0.05	Min. 0.05	Min. 0.05

Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

Manufacturing Standards

Manufacturing Standards	In compliance with ISO 9001 and ISO 13485
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only.



G·S·L

WWW.GLOVEMAN.CO.UK

Gloveman Supplies Ltd
 Vision House, Treleigh Ind Est,
 Redruth TR16 4AX
 t. 01209 314759
 f. 01209 314760
 sales@gloveman.co.uk